

Cleanroom Management International

We care for your compliance

Temperature calibrator TP 17650M TP Basic // Dry block // Room temperature...650 °C // RT...1202 °F



TP 17650M

TP 17650M - Highlights

- Version of our temperature calibrator TP 17650 specially optimised for the marine market
- Very easy operation with 4-button control and integrated reference temperature sensor
- Low weight and stable handle for easy transport
- Optional accessory: Transport bag or transport case with or without trolley
- Qualified for SIKA Gold Service

TP Basic

The temperature calibrators of the TP Basic series are characterised by their efficiency and portability. The series consists of dry block calibrators covering a wide temperature range and are used on site, e.g. in marine applications.

The easy operation, the integrated internal reference temperature sensor and the dry block calibration function ensure an extremely easy calibration process.

SIKA temperature calibrators

Temperature calibrators are used for the verification of the functionality and calibration of temperature measuring devices and temperature sensors. As the sole German manufacturer of these devices, we develop and produce our "Made in Germany" temperature calibrators with a special focus on long-term reliability and utmost accuracy in combination with easy operation. We can rely on more than 40 years of experience in doing this: SIKA's first dry block temperature calibrator was launched all the way back in 1980.

Every SIKA temperature calibrator is meticulously tested for accuracy and stability. This is attested by our standard calibration certificate, which we issue with every temperature calibrator, or by means of an optional DAkkS calibration certificate [German accreditation body]. This is to guarantee that you receive a perfect product which can be traced back to national and international temperature measurement standards.



Features

Easy operation

- The TP 17650M can be operated with only four buttons:
 Two arrow buttons for setting the target temperature,
 one button for confirmation and one return button
- Thus, temperatures can be set as easily as, for example, in the air conditioning system in your car
- Any operational errors can be nearly excluded. You do not need any specifically trained staff or time-consuming briefings



SIKA Gold Service

SIKA Gold Service provides a comprehensive service package for the regular recalibration of your temperature calibrator. You will benefit from exclusive savings and discounts as well as special promotions reserved to SIKA Gold Service members.

- You will save 33% in the recalibration of your temperature calibrator
- You will receive a 10% discount on any repairs that may become necessary
- You will receive preferential invitations to product presentations, symposia, practice days and exclusive training offers





Technical data

| TP 17650M | | | | |
|---------------------------------------|--|----------|--|--|
| Temperature range | Room temperature650 °C Room temperature1202 °F | | | |
| Dimension of the calibration insert | Ø 28 x 150 mm (calibration insert easily exchangeable) | | | |
| Dry block | | | | |
| Display accuracy | ±1 °C | ±1.8 °F | | |
| Temperature stability | ±0.1 °C | ±0.18 °F | | |
| Resolution of the temperature display | 1 °C 1 °F | | | |
| Reference temperature sensor | internal, fixed installation | | | |
| Dimensions | | | | |
| → Width → Height → Depth | 150 mm 330 + 70 mm (Handle) 270 mm | | | |
| Weight | Approx. 7.5 kg | | | |
| Power supply → Standard → Optional | 230240 VAC, 50 / 60 Hz 100115 VAC, 50 / 60 Hz 100240 VAC, 50 / 60 Hz | | | |
| Power consumption | Approx. 1000 W | | | |
| Display | | | | |
| Display | 2-line, 4-digit digital display red / green, unit °C / °F | | | |
| Approvals | | | | |
| | DIVICE COMM | | | |



Article numbers

To order a complete calibrator, you need two article numbers:

- 1. Calibrator
- 2. Calibration insert

In addition, depending on your individual calibration requirements, you can order additional calibration inserts, necessary certificates and other accessories.

| 1. Calibrator | | | | | |
|------------------------|-----------|-----------|-------------------------|--------------|----------------|
| Temperature range | | Function | Calibration insert [mm] | Power supply | Article number |
| Room temperature650 °C | RT1202 °F | Dry block | Ø 28 x 150 | 230 V | EP17650M281500 |
| Room temperature650 °C | RT1202 °F | Dry block | Ø 28 x 150 | 115 V | EP17650M281502 |
| Room temperature650 °C | RT1202 °F | Dry block | Ø 28 x 150 | 110240 V | EP17650M281503 |

| 2. Calibration insert | | | | |
|--|-----------|-------------------------|----------|----------------------|
| Bore holes [mm] | Function | Calibration insert [mm] | Material | Article number |
| 1x Ø 3.5, 1x Ø 6.5, 1x Ø 13.5 | Dry block | Ø 28 x 150 | Brass | EZ15028B03MS17 |
| 1x Ø 6.5 | Dry block | Ø 28 x 150 | Brass | EZ15028065MS00 |
| 2x Ø 3.5 | Dry block | Ø 28 x 150 | Brass | EZ15028B02MS09 |
| 1x Ø 3.5, 1x Ø 4.5 | Dry block | Ø 28 x 150 | Brass | EZ15028F02MS80 |
| 1x Ø 3.5, 1x Ø 6.5 | Dry block | Ø 28 x 150 | Brass | EZ15028H02MS01 |
| 1x Ø 3.5, 1x Ø 8.5 | Dry block | Ø 28 x 150 | Brass | EZ15028B02MS67 |
| 1x Ø 3.5, 1x Ø 6.5, 1x Ø 8.5, 1x Ø 10.5 | Dry block | Ø 28 x 150 | Brass | EZ15028C04MS15 |
| Without bore holes | Dry block | Ø 28 x 150 | Brass | EZ15028000MS00 |
| Calibration insert incl. 1 bore hole of choice | Dry block | Ø 28 x 150 | Brass | Please indicate bore |
| Each additional bore hole | Dry block | Ø 28 x 150 | Brass | holes in the order |

| 3. Calibration certificate - Select your calibration certificates as needed Each calibrator is already delivered with a standard calibration certificate (3 test points). | Article number |
|--|----------------|
| SIKA works calibration certificate (similar to standard calibration certificate + marking on the calibrator) | EKTPWP1FKT |
| DAkkS calibration certificate (3 test points + measurement uncertainty determination) | EKTPDAKKS1FKT |
| Each additional test point DAkkS calibration certificate | EKTPDAKKSZUSP |
| SIKA Gold Service works calibration certificate | EKTPGOLDWP |
| SIKA Gold Service DAkkS | EKTPGOLDDAKKS |

| 4. Accessories | Article number |
|--------------------------------|-----------------|
| Transport case without trolley | EZTPKOFFER020 |
| Transport case with trolley | EZTPKOFFER020TG |
| Transport bag | XE2193 |

CMI BELGIUM

Avenue de Ramelot 3 B - 1480 Saintes (Sint Renelde) Tel : +32 (0)2 531 11 21 info@cmitest.com

CMI THE NETHERLANDS

Luchthavenweg 81 NL – 5657 EA Eindhoven Tel : +31 (0)402 88 87 57 infonl@cmitest.com

CMI FRANCE

Ecoparc d'Affaires F- 41210 Neung-sur-Beuvron Tel : +33 (0)2 54 95 70 95 infofr@cmitest.com Agence Rhône Alpes : F- 26160 La Bégude-de-Mazenc Tel : +33 (0)4 75 54 57 26





Overview of SIKA temperature calibrators

| Temperature range (RT=Room temperature) | Function | Accuracy | | Features | Block dimensions [Ø mm x depth mm] | Туре |
|---|---|---|---|---|---------------------------------------|-------------|
| | Dry block | ±0.4 °C | ±0.72 °F | | 28 x 150 | TP 17200 |
| | Dry block | ±0.2 °C | ±0.36 °F | PC interface | 28 x 150 | TP 17200S |
| -55 °C 200 °C -67 °F 392 °F | Dry block | ±0.2 °C | ±0.36 °F | Touch screen PC interface External reference sensor Integrated measuring instrument | 28 x 150 | TP 37200E.2 |
| | Dry block | ±1 °C | ±1.80 °F | | 28 x 150 | TP 17165M |
| | Dry block | ±0.4 °C | ±0.72 °F | | 28 x 150 | TP 17165 |
| | Dry block | ±0.2 °C | ±0.36 °F | PC interface | 28 x 150 | TP 17165S |
| | Dry block | ±0.2 °C | ±0.36 °F | Touch screen PC interface External reference sensor Integrated measuring instrument | 28 x 150 | TP 37165E.2 |
| -35 °C 165 °C -31 °F 329 °F | Dry block | ±0.4 °C | ±0.72 °F | | 60 x 150 | TP 17166 |
| 01 1 027 1 | Dry block | ±0.2 °C | ±0.36 °F | PC interface | 60 x 150 | TP 17166S |
| | Calibration bath | ±0.1 °C | ±0.18 °F | PC interface | 60 x 170 | TP M165S |
| | Dry block Air Shield Insert Calibration bath Infrared Surface | ±0.3 °C ±0.099 °C ±0.1 °C ±0.5 °C ±1 °C | ±0.54 °F ±0.1782 °F ±0.18 °F ±0.9 °F ±1.88 °F | Touch screen PC interface External reference sensor Integrated measuring instrument | 60 x 170 | TP 3M165E.2 |
| -10 °C 100 °C 14 °F 212 °F | Dry block | ±0.05 °C | ±0.09 °F | PC interface | 7 x 6.5 x 150 | TP 17Zero |
| RT 200 °C RT 392 °F | Dry block | ±1 °C | ±1.80 °F | | 18 x 150 | TP 18200E |
| | Calibration bath | ±0.2 °C | ±0.36 °F | PC interface | 60 x 170 | TP M255S |
| RT 255 °C RT 491 °F | Dry block Calibration bath Infrared Surface | ±0.3 °C ±0.2 °C ±0.5 °C ±1°C | ±0.54 °F ±0.36 °F ±0.9 °F ±1.8 °F | Touch screen PC interface External reference sensor Integrated measuring instrument | 60 x 170 | TP 3M255E.2 |
| | Dry block | ±0.6 °C | ±1.08 °F | | 60 x 150 | TP 17450 |
| | Dry block | ±0.3 °C | ±0.54 °F | PC interface | 60 x 150 | TP 17450S |
| RT 450 °C RT 842 °F | Dry block Air Shield Insert Infrared Surface | ±0.3 °C ±0.2 °C ±0.5 °C ±1 °C | ±0.54 °F ±0.36 °F ±0.9 °F ±1.8 °F | Touchscreen PC interface External reference sensor Integrated measuring instrument | 60 x 150 | TP 37450E.2 |
| | Dry block | ±1 °C | ±1.80 °F | | 28 x 150 | TP 17650M |
| RT 650 °C RT 1202 °F | Dry block | ±0.8 °C | ±1.44 °F | | 28 x 150 | TP 17650 |
| | Dry block | ±0.4 °C | ±0.72 °F | PC interface | 28 x 150 | TP 17650S |
| RT 700 °C RT 1292 °F | Air Shield Insert | ±0.53 °C | ±0.954 °F | Touchscreen PC interface External reference sensor Integrated measuring instrument | 29 x 150 | TP 37700E.2 |
| RT 850 °C RT 1562 °F | Dry block | ±1 °C | ±1.80 °F | | 18 x 100 | TP 18850E |
| 400 °C 1300 °C 752 °F 2372 °F | Dry block | ±2°C | ±3.6 °F | PC interface | 28 x 200 | TP 281300E |

Subject to technical modifications and errors

